

The HiGen Series

Ordering Information

Herolab Cat.-No.	Type of Centrifuge	Short Description	Nominal Voltage	Phases	Hz	Accepted Voltage Range
21 10 100	<i>HiGen 21 C</i>	Three Phase Instrument for 50/60 cycles, Standard for Europe	400 V	3	50	340-440 V
			440 V	3	60	396-499 V
21 10 102	<i>HiGen 21 C</i>	Single Phase Instrument for 50 cycles, Standard for Europe	220-240 V	1	50	198-264 V
21 10 104	<i>HiGen 21 C</i>	Single Phase Instrument for 60 cycles	208-220 V	1	60	180-242 V
21 10 106	<i>HiGen 21 C</i>	Three Phase Instrument for 50/60 cycles	220 V	3	50	180-253 V
			220 V	3	60	187-264 V
21 10 108	<i>HiGen 21 C</i>	Single Phase Instrument for 50 cycles, Middle East Version	208-220 V	1	50	180-242 V
21 10 200	<i>HiGen XL</i>	Three Phase Instrument for 50/60 cycles, Standard for Europe	400 V	3	50	340-440 V
			440 V	3	60	396-499 V
21 10 202	<i>HiGen XL</i>	Single Phase Instrument for 50 cycles, Standard for Europe	220-240 V	1	50	198-264 V
21 10 204	<i>HiGen XL</i>	Single Phase Instrument for 60 cycles	208-220 V	1	60	180-242 V
21 10 206	<i>HiGen XL</i>	Three Phase Instrument for 50/60 cycles	220 V	3	50	180-253 V
			220 V	3	60	187-264 V
21 10 208	<i>HiGen XL</i>	Single Phase Instrument for 50 cycles, Middle East Version	208-220 V	1	50	180-242 V
21 10 900	Option Water-Cooling	Water-cooled refrigeration System for HiGen 21 C or HiGen XL				
21 10 650	<i>HiGen 18</i>	Single Phase Instrument for 50 cycles, Standard for Europe	220-240 V	1	50	198-264 V
21 10 652	<i>HiGen 18</i>	Single Phase Instrument for 50/60 cycles	100 V	1	50	90-110 V
			115 V	1	60	103-127 V
21 10 654	<i>HiGen 18</i>	Single Phase Instrument for 60 cycles	208-220 V	1	60	180-242 V
21 10 656	<i>HiGen 18</i>	Three Phase Instrument for 50/60 cycles	400 V	3	50	340-440 V
			440 V	3	60	396-499 V
21 10 658	<i>HiGen 18</i>	Single Phase Instrument for 50 cycles, Middle East Version	208-220 V	1	50	180-242 V
21 10 150	<i>HiGen F</i>	Single Phase Standard Instrument for 50/60 cycles	220-240 V	1	50/60	200-260 V
21 10 151	<i>HiGen F</i>	Single Phase Instrument for 50/60 cycles	115	1	50/60	100-130 V
21 10 050	<i>HiGen T</i>	Single Phase Standard Instrument for 50/60 cycles	220-240 V	1	50/60	200-260 V
21 10 052	<i>HiGen T</i>	Single Phase Instrument for 50/60 cycles	115 V	1	50/60	100-130 V

Herolab reserves the right to add or to change technical specifications without prior notice.